

Afinion AS100 Analyzer

Quiz

Name: _____

Date: _____

1. What is the operating temperature for the Afinion AS100 Analyzer?

- a. 25 – 50 F
- b. 49 – 79 F
- c. 59 – 89 F
- d. 65 – 95 F

2. True or False: The lid should always be manually closed?

3. How to properly open the lid:

- a. Manually lift open the lid



- d. b and c only
- e. All of the above

4. How often must the Afinion AS 100 Analyzer be cleaned?

- a. Daily
- b. Weekly
- c. Monthly
- d. Never

5. Which icon would you select to run a patient test?



6. When can a patient or control ID be entered?

- a. Before testing
- b. During testing
- c. After testing
- d. All of the above

7. True or False: Control testing mode is the default mode?

8. Which icon would you select to run a control test?



- d. All the above

9. When an operator ID  is entered it remains in use until:

- a. Another ID is entered
- b. It is manually deleted
- c. Analyzer is powered off
- d. All the above

10. What are the proper steps to perform prior to moving the Afinion AS100 Analyzer?

- a. Unplug and move
- b. Power down, unplug and move
- c. Move during the analysis of a test
- d. Never move the Afinion analyzer