Coag-Sense Test Strip Shipment QC Log Sheet and Test Log

Months: ____________________ Year: __________

Date Received	Date Tested	Tech	PT Test Strip Lot #	Low Control					High Control				
				Lot #	Expir Date	Expected Range	PT (sec)	P / F	Lot #	Expir Date	Expected Range	PT (sec)	P / F

INR TEST LOG

Date	ID #	PT	INR	Tech	Cleaned	# of Strips

Lab Director: ____ Date: _____________________________

Control strip lot numbers will be different than patient test strip and strip kit lot numbers. Run one high and one low control strip with each new lot of strips received and record the results. If your shipment includes multiple lot numbers, run controls for each lot received. Run controls when shipments arrive, as controls have limited shelf life. Mark "QC DONE" on the top of each box when completed. NWPC 2018